

**Notice of Allowability**

Application No.

10/630,706

Applicant(s)

MIKAMI ET AL.

Examiner

Art Unit

Regina Liang

2629

**-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--**

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to RCE filed 10/4/06.
2. ☒ The allowed claim(s) is/are 1-14.
3. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All b) ☐ Some\* c) ☐ None of the:
- ☐ Certified copies of the priority documents have been received.
  - ☒ Certified copies of the priority documents have been received in Application No. 10/083548.
  - ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\* Certified copies not received: \_\_\_\_\_.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.

**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
- (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
- 1) ☐ hereto or 2) ☐ to Paper No./Mail Date \_\_\_\_\_.
- (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

**Attachment(s)**

- ☐ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☒ Information Disclosure Statements (PTO/SB/08),  
Paper No./Mail Date 7-31-03
- ☐ Examiner's Comment Regarding Requirement for Deposit of Biological Material
- ☐ Notice of Informal Patent Application
- ☐ Interview Summary (PTO-413),  
Paper No./Mail Date \_\_\_\_\_
- ☒ Examiner's Amendment/Comment
- ☐ Examiner's Statement of Reasons for Allowance
- ☐ Other \_\_\_\_\_

  
Regina Liang  
Primary Examiner  
Art Unit: 2629

### EXAMINER'S AMENDMENT

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Mr. Leonid Thenor on 11/20/06.

The application has been amended as follows:

As to claim 6, line 15, change "said sampling period" to – a sampling period--; line 16, after "switch" insert --; and subsequently applied said held voltage signal to said driving elements by opening said sampling switch and closing said driving switch--; line 17, change "a sampling period" to – said sampling period--.


As to claim 11, line 17, change "said sampling period" to – a sampling period--; line 18, after "switch" insert --; and subsequently applied said held voltage signal to said driving elements by opening said sampling switch and closing said driving switch--; line 19, change "a sampling period" to – said sampling period--.

2. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Regina Liang whose telephone number is (571) 272-7693. The examiner can normally be reached on Monday-Friday from 8AM to 5:00PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Richard Hjerpe, can be reached on (571) 272-7691. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 2629

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

  
Regina Liang  
Primary Examiner  
Art Unit 2674

11/20/06